

기술 자료

# Fluke BP291 4800 mAh High Capacity Li-Ion Battery for Fluke 190-Series-II



## 주요 기능

- 4800 mAh rechargeable Li-ion battery
- Compatible with ScopeMeter® 190 Series III, 190 Series II, MDA-500 Series Motor Drive Analyzers and 430-II Series Power Quality Analyzers

## 제품 개요: Fluke BP291 4800 mAh High Capacity Li-Ion Battery for Fluke 190-Series-II

The Fluke BP291 Li-ion Battery is a 4800 mAh high-capacity battery for use in Fluke 190 Series II and Series III ScopeMeter Test Tools, MDA-500 Series III Motor Drive Analyzers and 430-II Series Power Quality Analyzers. BP291 High-Capacity Li-ion batteries include a charge-status indicator so you can quickly and easily verify charge status in the field, and may be charged either in the instrument itself, or using the option [EBC290 External Charger](#).

## 제품 사양: Fluke BP291 4800 mAh High Capacity Li-Ion Battery for Fluke 190-Series-II

Specifications\*

Battery	Li-ion, 10.8 V, 5000 mAh, 54 Wh configuration 3ICR19/66-2
Safety	Pollution degree 2
Operating temperature	0 °C to +50 °C (charging may slow or stop at environmental temperatures >40 °C)
Storage temperature	-20 °C to +60 °C
Recommended storage temperature	<35 °C
Humidity	0 °C to 10 °C noncondensing 10 °C to 30 °C @ 95 % (±5 %) rh 30 °C to 40 °C @ 75 % (±5 %) rh
Altitude	3000 m
Weight	330 g
Dimensions	22 mm x 150 mm x 59 mm

## 모델



### **Fluke BP291**

4800 mAh High capacity Li-ion battery for Fluke 190 seriesII

---



**Fluke. Keeping your world up and running.®**

Fluke Corporation  
PO Box 9090, Everett, WA 98206 U.S.A.

(주)한국플루크Fluke Korea  
Tel.02.539.6311  
(주)한국플루크 대구지사  
Tel.053.382.6311  
www.fluke.co.kr

Fluke Korea  
서울특별시 강남구 영동대로 517,10층 1002호  
(삼성동, 아셈타워)  
©2024 Fluke Corporation.  
사양은 예고 없이 변경될 수 있습니다.  
01/2024

이 문서의 수정은 Fluke Corporation 의 서면 허가 없이는 허용되지 않습니다.